Arlington, TX 76012 Pocomoke City, MD 21851 Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

# **Product Specification**

VeriSpec® Succinate - OOC(CH<sub>2</sub>)<sub>2</sub>COO<sub>2</sub> Standard for Ion Chromatography 1000 ppm in H<sub>2</sub>O Manufactured and Tested in an ISO 17025/Guide 34 Facility

Product Number: RV010942 Lot Number: SAMPLE

Manufacture Date: N/A Expiration Date: N/A

Component	Certified Value	Uncertainty	Starting Material*	Traceability
OOC(CH <sub>2</sub> ) <sub>2</sub> COO <sub>2</sub>	1000 ppm <sup>(b)</sup>		$C_{12}H_{22}O_{4}$ *	

<sup>\*</sup> Starting material purity is not a certified value.

Matrix: H\_2O

### Method(s) of certification used:

The certified value was obtained using Ion Chromatography calibration

# Concept of certification and traceability statement:

This certified reference material is produced using a high-purity starting material, acid from sub-boiling and 18 MOhm deionized water. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02

Property of the result of a measurement whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties (ISO VIM)

The metrological traceability is assured through calibration on Ion Chromatographs. The calibration curve is drawn using a series of standard solutions prepared from a certified reference material traceable to SI of NIST (SRM) and of accredited according to ISO/IEC 17025 and/or ISO Guide 34 laboratories/producers. All contributions in relation to the certification of standard solutions are considered when evaluating the uncertainty. The measurement results are traceable to SI. All analytical balances used for the preparation of the solution are calibrated yearly under an in-house procedure with analytical weights, traceable to DKD and are daily checked. Class A laboratory glassware is used.

The results from temperature measurement are traceable to SI. The thermometers used for solution's calibration are calibrated from an ISO 17025 accredited laboratory. The ambient conditions are controlled with a hygrometer calibrated from an ISO 17025 accredited laboratory.

#### Level of homogeneity:

The solution was mixed according to an in-house procedure and is guaranteed to be homogeneous.

To ensure sufficient homogeneity of the sample prior to use thoroughly mix by inversion.

### Intended use:

For Laboratory Use Only This CRM is intended for:

> Calibration of Ion Chromatographs Validation of analytical methods Preparation of "working reference samples"

Detection limit and linearity studies

This statement is not intended to restrict the use for other purposes.

#### Instructions for the correct use of this reference material:

This certified reference material can be used directly or can be diluted in an appropriate high-purity matrix. Only clean class A glassware should be used. Do not pipet from container. Obtained concentration (in mg/l) after dilution is a result from the multiplication of certified value of CRM concentration and the CRM's volume used for dilution and divided into the flask's volume used for dilution.

#### Hazardous situation:

The normal laboratory safety precautions should be observed when working with this RM. Further details for the handling of this RM are available as safety data sheet.

Lot Number: SAMPLE Product Number: RV010942 Page 1 of 2 Version: 1.0

## Stability and storage:

This CRM is with a guaranteed stability until  $\pm 0.5\%$  of the certified concentration within its shelf-life. Stability is guaranteed provided that the solution is kept in its original packaging, tightly closed under normal laboratory conditions.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RV010942-100C	100 mL amber glass	24 months
RV010942-250C	250 mL amber glass	24 months
RV010942-500C	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This Certified Reference material was produced under a quality management system that is accredited to ISO/IEC 17025 and ISO Guide 34.

Version: 1.0 Lot Number: SAMPLE Product Number: RV010942 Page 2 of 2