

Product Specification

Acetic Acid, 20% (v/v)

Lot Number: SAMPLE

Product Number: 138

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Acetic Acid	64-19-7	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Sodium Hydroxide/Thymol Blue)	19.94-20.06 % (v/v)	N/A	84

Specification	Reference
Acetic acid, 20%	TAPPI (T 611 cm-97)
Acetic acid solution, 20% solution	TAPPI (T 622 cm-84)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
138-2.5	10 L Cubitainer®	36 months
138-32	1 L natural poly	36 months
138-5	20 L Cubitainer®	36 months
138-1	4 L natural poly	36 months
138-16	500 mL natural poly	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."