

Product Specification

Bromine Number Titration Solvent, Dichloromethane Formulation

Lot Number: SAMPLE

Product Number: 1191

Manufacture Date: N/A

Expiration Date: N/A

Dichloromethane must be used in place of 1,1,1-Trichloroethane for all operations of this ASTM procedure (i.e., for the dilution of samples and as the titration blank) when this reformulated Titration Solvent is used.

Name	CAS#	Grade
Acetic Acid	64-19-7	ACS
Methylene Chloride	75-09-2	ACS
Methyl Alcohol	67-56-1	ACS
Sulfuric Acid	7664-93-9	ACS
Water	7732-18-5	ACS/ASTM/USP/EP

Test	Specification	Result
Appearance	Colorless liquid	N/A

Specification	Reference
---------------	-----------

Test Method for Bromine Numbers of Petroleum Distillates and Commercial Aliphatic Olefins by Electrometric Titration ASTM (D 1159)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1191-32	1 L amber glass	24 months
1191-5PT	2.5 L safety-coated glass	24 months
1191-1	4 L safety-coated glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."