

## **Product Specification**

## Bromocresol Green Indicator, 1% (w/v)

Manufacture Date: N/A

Lot Number: SAMPLE

## Product Number: R1209000

Expiration Date: N/A

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Bromocresol Green, Sodium Salt	62625-32-5	ACS	
Test	Specification	Result	
Appearance	Dark blue liquid	N/A	
Suitability for Use	Blue (pH >5.4) - Green (pH 4.5) - N/A Vollow (pH $\leq$ 4.0)		

Yellow (pH < 4.0)Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with  $\varepsilon$ thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)	
R1209000-1A	1 L natural poly	24 months	
R1209000-500A	500 mL natural poly	24 months	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."