

Product Specification

Bromophenol Blue Indicator, 0.1% (w/v)

Lot Number: SAMPLE

Product Number: 1353

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Bromophenol Blue, Sodium Salt	62625-28-9	ACS

Test	Specification	Result
Appearance	Dark reddish-purple liquid	N/A
Suitability for Use	Yellow (pH < 3.0) - Green (pH 3.4 to 3.7) - Blue (pH > 4.6)	N/A

Specification	Reference
Bromophenol Blue Indicator Solution (1 g/L)	ASTM (D 3869 B)
Bromophenol blue indicator solution, pH 3.7 indicator	APHA (2310 B)
Bromophenol blue indicator	TAPPI (T 249 cm-85)
Bromophenol blue, aqueous	TAPPI (T 610 sp-97)
Bromophenol Blue Indicator Reagent Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1353-4	120 mL natural poly	24 months
1353-32	1 L natural poly	24 months
1353-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."