

Product Specification

Buffer, Reference Standard, pH 3.56 ± 0.01 at 25°C

Lot Number: SAMPLE

Product Number: 1498

Manufacture Date: N/A

Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
pH at 25°C	3.550-3.570	N/A	185 & 186
pH at 25°C Uncertainty	0.01	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1498-2.5	10 L Cubitainer®	24 months
1498-32	1 L natural poly	24 months
1498-5	20 L Cubitainer®	24 months
1498-1	4 L natural poly	24 months
1498-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



**LABORATORY
ACCREDITATION
BUREAU** a division of **A-S-B**
ACCREDITED ISO/IEC 17025