

## **Product Specification**

Manufacture Date: N/A

Expiration Date: N/A

## Buffer, Reference Standard, pH $12.00 \pm 0.01$ at $25^{\circ}C$

## Lot Number: SAMPLE

Product Number: 1615

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	15	20	25	30	35	40
pН	12.35	12.17	11.99	11.78	11.62	11.46

CAS#	Grade	
7732-18-5	ACS/ASTM/USP/EP	
7447-40-7	ACS	
1310-73-2	Reagent	
	7732-18-5 7447-40-7	7732-18-5 ACS/ASTM/USP/EP   7447-40-7 ACS

Test	Specification	$\mathbf{Result}$	NIST SRM#
Appearance	Colorless liquid	N/A	
pH at 25°C	11.990-12.010	N/A	186 & 191
pH at 25°C Uncertainty	0.02	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with  $\varepsilon$  thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)	
1615-2.5 10 L Cubitainer®		18 months	
1615-32 1 L natural poly		18 months	
1615-20B	200 x 20 mL pack	18 months	
1615-5	20 L Cubitainer®	18 months	
R1615000-20F1	20 mL pouch	18 months	
1615-20	20 x 20 mL pack	18 months	
1615-1	4 L natural poly	18 months	
1615-1CT	4 L Cubitainer®	18 months	
1615-16	500 mL clear PET-G	18 months	

Recommended Storage: 15°C - 30°C (59°F - 86°F)



This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."