

Product Specification

Ceric Ammonium Nitrate, 0.100 Normal in 34.5% (v/v) Perchloric Acid

Lot Number: SAMPLE**Product Number:** R1886000**Manufacture Date:** N/A**Expiration Date:** N/A

This solution MUST be restandardized every 15 days according to Stone AN250L.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Perchloric Acid	7601-90-3	ACS
Ceric Ammonium Nitrate	16774-21-3	ACS

Test	Specification	Result	NIST SRM#
Appearance	Orange liquid	N/A	
Assay (vs. Sodium Oxalate)	0.0995-0.1005 N at 20°C	N/A	40

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R1886000-500C	500 mL amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."