

Product Specification

Decolorized Pararosaniline Solution R
EP Reagent 1062201**Lot Number:** SAMPLE**Product Number:** R2545500**Manufacture Date:** N/A**Expiration Date:** N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Metabisulfite	7681-57-4	EP
Basic Fuchsin	569-61-9	Certified
Charcoal, Activated	7440-44-0	USP
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result
Appearance	Colorless to slightly pink liquid	N/A
Sensitivity to Aldehyde	To Pass Test	N/A

Specification	Reference
Decolorized Pararosaniline Solution R	EP (1062201)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R2545500-500C	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)