

Product Specification

Dichlorofluorescein TS, 0.1% (w/v) in Dilute Alcohol

Lot Number: SAMPLE

Product Number: 2549

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Hydroxide	1310-73-2	Reagent
Methyl Alcohol	67-56-1	Reagent
Dichlorofluorescein	76-54-0	ACS

Test	Specification	Result
Appearance	Greenish-yellow liquid	N/A
Suitability for Use	Adsorption: greenish-yellow (excess Chloride) - Rose (excess Silver ions)	N/A

Specification	Reference
Dichlorofluorescein TS	USP (N/A)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2549-4	120 mL natural poly	12 months
2549-16	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)