

# Product Specification

## Iodine Monochloride Solution, Wijs, for Iodine Absorption Number of Fats and Oils

Lot Number: SAMPLE

Product Number: 4110

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Acetic Acid	64-19-7	ACS
Iodine Monochloride	7790-99-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	N/A	
Iodine/Chlorine Ratio	1.05-1.15	N/A	
Total Halogen Content (vs. Sodium Thiosulfate/Starch)	0.189-0.193 N at 20°C	N/A	136

Specification	Reference
Wijs Iodine Solution	AOAC (920.159)
Iodochloride TS	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4110-100	100 mL amber glass	12 months
4110-32	1 L amber glass	12 months
4110-5PT	2.5 L safety-coated glass	12 months
4110-1	4 L amber glass	12 months
4110-1SC	4 L safety-coated glass	12 months
4110-16	500 mL amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."