

Product Specification

Iodine and Potassium Iodide TS 1

Lot Number: SAMPLE

Product Number: 4124

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	N/A	
Assay (vs. Sodium Thiosulfate/Starch)	0.1574-0.1578 N at 20°C	N/A	136

Specification	Reference
---------------	-----------

Iodine and Potassium Iodide TS 1

USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4124-32	1 L amber glass	12 months
4124-8	250 mL amber glass	12 months
4124-16	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)