

Product Specification

Lead Acetate TS, 2% (w/v) Alcoholic Solution

Lot Number: SAMPLE

Product Number: 4293.8

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Methyl Alcohol	67-56-1	Reagent
Lead Acetate Trihydrate	6080-56-4	ACS

Test	Specification	Result
Appearance	Colorless liquid	N/A

Specification	Reference
Lead Acetate TS, Alcoholic	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4293.8-32	1 L natural poly	12 months
4293.8-16	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."