Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

PAN Indicator, 0.1% (w/v) in Methanol Indicator for Complexometric Titrations

Lot Number: SAMPLE Product Number: R5487100

Manufacture Date: N/A
Expiration Date: N/A

Arlington, TX 76012

Name	CAS#	Grade	
Methyl Alcohol	67-56-1	ACS	
1,2-Pyridylazo-2-naphthol	85-85-8	ACS	

Test	Specification	Result
Appearance	Amber liquid	N/A
Suitability for Use	Deep Red (Uncomplexed Copper Ions) - Yellow (Excess EDTA)	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ε thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5487100-100C	100 mL amber glass	24 months
R5487100-500C	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.0 Lot Number: SAMPLE Product Number: R5487100 Page 1 of 1