

Product Specification

. . NT/ A

Potassium Dichromate, 4% (w/v)

Lot Number: SAMPLE	Product Number: R6042800		Manufacture Date: N/A Expiration Date: N/A
Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Potassium Dichromate	7778-50-9	ACS	
Test	Specification	Result	NIST SRM#
Appearance	Orange liquid	N/A	
Assay (vs. Sodium Thiosulfate/Starch)	3.96-4.04 % (w/v)	N/A	136
Specification	Reference		
Potassium Dichromate, 4% (w/v)	TAPPI (T421 om-97)		

Potassium Dichromate, 4% (w/v) Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ε thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validatec methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R6042800-500A	500 mL natural poly	24 months
,		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."