Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Potassium Fluoride, 10% (w/v)

Lot Number: SAMPLE Product Number: R6098000

Manufacture Date: N/A Expiration Date: N/A

Arlington, TX 76012

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Fluoride	7789-23-3	ACS

Test	Specification	Result
Appearance	Colorless liquid	N/A
Specific Gravity	1.07-1.09 at 25°C	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ϵ thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R6098000-1A	1 L natural poly	24 months
R6098000-4A	4 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: R6098000 Page 1 of 1