Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Potassium Phosphate Monobasic, 176.5 g/L

Lot Number: SAMPLE Product Number: 6451

Manufacture Date: N/A Expiration Date: N/A

Arlington, TX 76012

Pocomoke City, MD 21851

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Potassium Dihydrogen Phosphate	7778-77-0	ACS	

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Sodium Hydroxide)	171.5-181.5 g/L	N/A	84

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ϵ thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number Size / Package Ty		Size / Package Type	Type Shelf Life (Unopened Container)	
	6451-32	1 L natural poly	12 months	
	6451-30	30 gal drum	12 months	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials \cdots Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: 6451 Page 1 of 1