

Product Specification

Silver Nitrate-Nitric Acid Reagent, for Sulfate Determination

Lot Number: SAMPLE

Product Number: 7045

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Silver Nitrate	7761-88-8	ACS
Nitric Acid	7697-37-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Potassium Thiocyanate/Volhard's Indicator)	0.095-0.105 N AgNO ₃	N/A	999

Specification	Reference
Silver Nitrate-Nitric Acid Reagent	APHA (4500-SO42- D)
Silver Nitrate-Nitric Acid Reagent	APHA (4500-SO42- C)
Silver Nitrate-Nitric Acid Reagent	EPA (375.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7045-4	120 mL amber glass	24 months
7045-16	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)