

Product Specification

Sodium Carbonate, 1.00 Normal

Lot Number: SAMPLE

Product Number: 7190

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Sulfuric Acid/Methyl Orange)	0.999-1.001 N at 20°C	N/A	723

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7190-2.5	10 L Cubitainer®	18 months
7190-32	1 L natural poly	18 months
7190-5	20 L Cubitainer®	18 months
7190-1	4 L natural poly	18 months
7190-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)