

Product Specification

Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Stannous Chloride Reagent II for Phosphorus (Phosphate) Analysis (with extraction)

Lot Number: SAMPLE Product Number: 7998

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade	
Glycerin	56-81-5	ACS	
Stannous Chloride Dihydrate	10025-69-1	ACS	

Test	Specification	Result
Appearance	Colorless to light tan liquid	N/A

Specification	Reference
	 A DITA (4500 B B)

Dilute Stannous Chloride Reagent II

APHA (4500-P D)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7998-4	120 mL natural poly	12 months
7998-16	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: 7998 Page 1 of 1