

Product Specification

Sulfo Orange Indicator, 0.04% (w/v)

Lot Number: SAMPLE

Product Number: 8113

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfo Orange	547-57-9	Indicator

Test	Specification	Result
Appearance	Reddish-orange liquid	N/A
Suitability for Use	Yellow (pH < 11.0) - Orange (pH > 12.6)	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8113-4	120 mL amber glass	24 months
8113-32	1 L amber glass	24 months
8113-5	20 L Cubitainer®	24 months
8113-16	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)