

Product Specification

Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Sulfuric Acid, Dilute R (98 g/L H₂SO₄)

Lot Number: SAMPLE Product Number: 8140

Manufacture Date: N/A Expiration Date: N/A

| Name | CAS# | Grade |
|---------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sulfuric Acid | 7664-93-9 | ACS |

| Test | Specification | \mathbf{Result} | NIST SRM# |
|--|-----------------------|-------------------|-----------|
| Appearance | Colorless liquid | N/A | |
| Assay (vs. Sodium Hydroxide/Phenolphthalein) | 1.995-2.005 N at 20°C | N/A | 84 |

| Specification | Reference |
|--------------------------|--------------|
| Sulphuric acid, dilute R | EP (1086804) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ε thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 8140-4 | 120 mL natural poly | 36 months |
| 8140-8 | 250 mL natural poly | 36 months |
| 8140-16 | 500 mL natural poly | 36 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: 8140 Page 1 of 1