

Product Specification

Trichloroacetic Acid, 80% (w/v)

Lot Number: SAMPLE

Product Number: 8693

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Trichloroacetic acid	76-03-9	ACS
Water	7732-18-5	ACS/ASTM/USP/EP
Charcoal, Activated	7440-44-0	USP

Test	Specification	Result	NIST SRM#
Appearance	Colorless to tan liquid	N/A	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	79.6-80.4 % (w/v)	N/A	84

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8693-4	120 mL amber glass	12 months
8693-32	1 L amber glass	12 months
8693-16	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)