

Product Specification

Zinc Sulfate, 0.0500 Molar (M/20)

Lot Number: SAMPLE

Product Number: 9660

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Zinc Sulfate Heptahydrate	7446-20-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. EDTA/Xylenol Orange)	0.04999-0.05001 Molar	N/A	915

Specification	Reference
Zinc Sulfate, Twentieth-Molar (0.05 M) Volumetric Solution	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
9660-32	1 L natural poly	18 months
9660-5	20 L Cubitainer®	18 months
9660-8	250 mL natural poly	18 months
9660-1	4 L natural poly	18 months
9660-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."