

Product Specification

Zirconyl-SPADNS Reagent, for Fluoride Analysis

Lot Number: SAMPLE

Product Number: 9850

Manufacture Date: N/A

Expiration Date: N/A

The SPADNS Reference Solution (RICCA CHEMICAL COMPANY Cat. No. 7990) is also available for use to set the photometer reference point to zero absorbance before measuring the absorbance of appropriate Fluoride Standards (for standard curve preparation) and the absorbance of Fluoride test samples.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS
Zirconyl Chloride Octahydrate	13520-92-8	Practical
Sodium 2-(parasulfophenylazo)-1,8-dihydroxy-3,6-naphthalene Disulfonate	23647-14-5	Indicator

Test	Specification	Result
Appearance	Dark red liquid	N/A

Specification	Reference
Acid Zirconyl-SPADNS Reagent	APHA (4500-F- D)
Acid-Zirconyl-SPADNS Reagent	EPA (340.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
9850-100	100 mL amber glass	24 months
9850-32	1 L amber glass	24 months
9850-16	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."