

Product Specification

Water, ACS Reagent Grade, ASTM Type I, ASTM Type II
Packaged in glass containers

Lot Number: SAMPLE

Product Number: 9151

Date Packaged: N/A

Expiration Date: N/A

Packaged in glass for applications requiring extremely low levels of organics. This product is manufactured by a continuous process that has been validated to produce high purity Water meeting, as a minimum, the specifications below at the time of packaging. Although this water is not distilled it has a chemical purity equal to or better than double distilled water. The analysis results are from the most recent scheduled monitoring. This product also meets ASTM Type III and Type IV specifications.

Contains

H₂O CAS No. 7732-18-5 MW 18.02

Test	Specification	Result
Appearance	Colorless liquid	N/A
Chloride (Cl)	max 1 ppb	N/A
Conductivity at 25°C	max 0.056 µS/cm	N/A
Heavy Metals (as Pb)	max 0.01 ppm	N/A
Nitrate (NO ₃)	max 0.4 ppm	N/A
Organic Carbon (TOC)	max 50 ppb	N/A
Oxidizable Substances (Permanganate Retention)	To Pass Test	N/A
Phosphate (PO ₄)	max 1.0 ppm	N/A
Silicate (as SiO ₂)	max 3 ppb	N/A
Sodium (Na)	max 1 ppb	N/A
Sulfate (SO ₄)	max 1.0 ppm	N/A

Specification	Reference
ASTM Type II Water	EPA (SW-846) (7197)
ASTM Type II Water	EPA (SW-846) (7195)
Deionized, distilled water (ASTM Type II)	EPA (200.7)
Reagent Water	ACS (N/A)
Reagent Water	EPA (SW-846) (N/A)
Reagent Water, Type I	ASTM (D 1193)
Reagent Water, Type II	ASTM (D 1193)
Reagent Water, Type III	ASTM (D 1193)
Reagent Water, Type IV	ASTM (D 1193)

This water is specially purified by conventional water softening, prefiltration, activated carbon organic adsorption, reverse osmosis, mixed bed triple deionization, ultraviolet light irradiation, and 0.2 micron membrane filtration. Consult RICCA CHEMICAL COMPANY Technical Reference Document TRD02 (for lot numbers starting with 1) or TRD03 (for lot numbers starting with 2) for details of the water purification process.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
9151-1	4 L amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."